



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0427-17

ITEM DESCRIPTION- Battery Testing System

Refer our Press Tender Notice No. IISERT/S&P/09/17 dated 12.01.2018 for procurement of Battery Testing System. Tender Reference Number - IISERT-PUR-0427-17.

Pre-Bid meeting was held on January 18th, 2018 at 16.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders if any and replied thereto as detailed if required in Annexure -II

The representatives present were satisfied if any with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iisertirupati.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

23.1.2018

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF BATTERY TESTING SYSTEM
TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0427-17

DATE: 18.01.18

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No. 22, Point No. 1. Number of Channels.	<p>Modified the following points,</p> <p>Page No. 22, Point No. 1. "minimum 60" modified to "60 - 65".</p> <p>Page No. 22, Point No. 3. "55" modified to "55 - 60" and "remaining channels" modified to "remaining channels (5 to 10)"</p> <p style="padding-left: 40px;">"Channels 1-55" modified to "1 to 55 - 60".</p> <p style="padding-left: 40px;">"Remaining Channels" modified to "Channels 5 to 10".</p> <p style="padding-left: 40px;">"EIS is required for all the 60 channels" modified to "EIS is required for all the 60-65 channels"</p> <p style="padding-left: 40px;">Point No. 4. "60 channels" modified to "60-65 channels".</p>



IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF BATTERY TESTING SYSTEM
COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0427-17

DATE: 18.01.18

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----