



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0163-17

ITEM DESCRIPTION- SOLUTION IR SPECTROPHOTOMETER

Refer our Press Tender Notice No.IISERT/S&P/04/17 dated 13.10.2017 for procurement of Solution IR Spectrophotometer. Tender Reference Number - IISERT-PUR-0163-17.

Pre-Bid meeting was held on October 24th , 2017 at 14.30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

03.11.2017

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF SOLUTION IR SPECTROPHOTOMETER

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0163-17

DATE : 03.11.2017

S.No	Query/Clarification Sought	Clarification / Amendment
01	Change in ATR/Transmittance Element in probe	We need a probe with adjustable-path length to follow samples at low concentration. Therefore, no change in the Tender specifications
02	Change in spectral range for probe to 3300 to 900 cm^{-1}	The spectral range under the performance in optical system is changed to 3300 to 900 cm^{-1} (with ZnSe beam splitter)



IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF SOLUTION IR SPECTROPHOTOMETER

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0163-17

DATE : 03.11.2017

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----