



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH  
TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0120-16

ITEM DESCRIPTION- MICROWAVE BASED MANUAL PEPTIDE SYNTHESIS

Refer our Press Tender Notice No.IISER/S&P/09/16 dated 7.10.2016 for procurement of Microwave Based Manual Peptide Synthesis. Tender Reference Number - IISERT-PUR-0120-16

Pre-Bid meeting was held on October 20<sup>th</sup> , 2016 at 14.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Pune and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iiserpune.ac .in](http://www.iiserpune.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

20.10.2016

Sd/-  
Assistant Registrar (S&P)

IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF MICROWAVE BASED MANUAL PEPTIDE SYNTHESIS**

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER – IISERT-PUR-0120-16

DATE : 24.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
01	<p>Page No - 22, Chapter 4, Point no. 1c</p> <p>System should be able to conduct sub ambient temperature reaction from -80 Deg C to +300 deg C, suitable accessory to be offered in optional.</p> <p>Query by one of the bidders “What is the temperature range required for the intended use of the instrument?”</p>	<p>Amended Specifications:</p> <p>For intended use of the instrument, a temperature range of -10 DegC to + 200 deg C is required.</p>
02	<p>Page No -23, Chapter 4, Point no. 4</p> <p>Query by one of the bidders “How many simultaneous reactions the instrument needs to perform?”</p>	<p>For the current and future use of the instrument, one reaction at a time is only required. Capability to perform multiple simultaneous reactions will also comply.</p>



IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF MICROWAVE BASED MANUAL PEPTIDE SYNTHESIS**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER – IISERT-PUR-0120-16

DATE : 24.10.16

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----