



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

Transit campus: C/o Sree Rama Engineering College Campus, Rami Reddy Nagar,
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: www.iisertirupati.ac.in

CORRIGENDUM: TECHNICAL BID

Tender Id	2021_ISRTP_625939_1
Tender Reference Number	IISERT/PUR/0182/20
Tender Title	Supply, Installation and Commissioning of Fluorimeter

With reference to the above tender the Corrigendum has been made due to change in technical specification.

Sr.No	Existing Specification	Amendment Specification
01	Point no. 2: Detectors	The specification has been modified as Detectors with spectral coverage in wavelength range $\lambda \leq 850\text{nm}$ (steady state Fluorescence and phosphorescence decay lifetime) and $\lambda \leq 1550\text{nm}$ (Emission and Anisotropy). The performance of detectors in these ranges must be supported by publications
02	Point no. 6 (Additional accessories): -Cuvettes	Fluorescence Quartz Cuvettes 1. 5 No's with path length 3 mm or 10 mm with volumes: $\leq 50 \mu\text{l}$; $\leq 100 \mu\text{l}$; $\leq 500 \mu\text{l}$. <u>Optional accessories</u> Quartz Cuvettes (5 No's) with Path length-2 mm of similar volumes

All the Prospective Bidders are required to take cognizance of the proceedings of the document before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER Tirupati website <http://www.iisertirupati.ac.in> will remain unchanged.

Assistant Registrar (Admin & Purchase)