



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0054/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF UV-VISIBLE SPECTROPHOTOMETER

Refer our Tender No: IISERT/PUR/0054/20 dated 27-April-2020 Supply, Installation and Commissioning of UV-Visible Spectrophotometer

Pre-Bid meeting was held on May 05th, 2020 at 17:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

05.05.2020

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF UV-VISIBLE SPECTROPHOTOMETER
TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0054/20

DATE: 05/05/2020

S.No	Query/Clarification Sought	Clarification / Amendment
1.	<u>Wavelength range:</u> -Most of the UV-VIS Spectro. Provides only up to 900nm When considered variable band width & the entire UV-Vis range covers within 185 to 900 nm. Can be it changed to 900 nm since all major applications related to UV/Vis range can be done within 900 nm range.	<u>Wavelength range:</u> -We have many compounds having absorption peaks at 700 nm or more and their end absorbance is usually go to 1000 or more, hence required almost 1100 nm. -However, It can be changed to 900 nm or better for your requirement.
2.	<u>Photometric noise:</u> < 0.0001 Abs. Can it be changed to <0.0002 Abs or better?	<u>Wavelength range:</u> - We need instrument with higher noise performance, we cannot the change specification.
3.	<u>Temperature controller:</u> Other accessories: 89054A Thermostatted cell holder. A jacketed cell holder for standard 1 cm cuvettes with provision for attaching to a water bath for temperature control.	<u>Temperature controller:</u> Other accessories: "89054A" can be deleted. Any thermostatted jacketed cell holder for standard 1 cm cuvettes with provision for attaching to a water bath for temperature control.



IISER TIRUPATI

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF UV-VISIBLE SPECTROPHOTOMETER
COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0054/20

DATE: 05/05/2020

Sr.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----