



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER – Tender No: : IISERT/PUR/0059/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING BATTERY TESTER

Refer our Tender No.IISERT/PUR/0059/20 dated 27/04/2020 for procurement of Supply, Installation and Commissioning of Battery Tester

Pre-Bid meeting was held on May 05<sup>th</sup>, 2020 at 14:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

05.05.2020

Sd/-  
Assistant Registrar (S&P)



**IISER TIRUPATI**  
**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF BATTERY TESTER**  
**TECHNICAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0059/20**

**DATE: 05/05/2020**

<b>S.No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
1.	<b><u>Testing current rate:</u></b> Reduction in the Testing current rate from $\pm 200$ mA or higher to $\pm 100$ mA	We are not only using this tester for battery materials, also the same tester will be used for Supercapacitor materials as well. So, we require a high current rating specified in the Tender document. Therefore, altering the current rate is not possible which limits our testing capability.
2.	<b><u>Measurement Resolution</u></b> 16 bit & above	One of the companies has a 32-bit resolution which is better than our requirement. Moreover, this does not require any modification, since we already mentioned "16 bit & above".



ANNEXURE -III

**IISER TIRUPATI**

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF BATTERY TESTER  
COMMERCIAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0059/20**

**DATE: 05/05/2020**

<b>S.No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
1	Nil	Nil