



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0071/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF MAGNETIC STIRRER WITH HOTPLATE, HEAT SENSOR AND LCD DISPLAY

Refer our Tender No: IISERT/PUR/0071/20 dated 19-May-2020 for procurement of Supply, Installation and Commissioning of Magnetic stirrer with hotplate, heat sensor and LCD display.

Pre-Bid meeting was held on May 27th, 2020 at 14:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

27.05.2020

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF MAGNETIC STIRRER WITH HOTPLATE, HEAT SENSOR
AND LCD DISPLAY
TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0071/20

DATE: 27/05/2020

S.No	Query/Clarification Sought	Clarification / Amendment
1.	-Hot plate diameter: 140 mm to 150 mm. Can it be 135mm?	No, We cannot change it, as we have specific heat-on-attachments.
2.	-Temperature accuracy hotplate: ± 1 Can it be ± 2	No, The temperature accuracy of hotplate cannot be to ± 2 °C.
3.	-Display: LCD Display for easy value setting Can it be a LED display also?	Yes, Display can also be allowed in LED mode.
4.	-Heating Power: 800W or more Can it be 630W or more?	Yes, Heating Power can be change to 630W or more.



ANNEXURE -III

IISER TIRUPATI

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF MAGNETIC STIRRER WITH HOTPLATE, HEAT SENSOR
AND LCD DISPLAY
COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0071/20

DATE: 27/05/2020

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----