



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT-PUR-0308-20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF HPC.

REFER OUR TENDER NO: **IISERT-PUR-0308-20** DATED: 09/04/2021 FOR SUPPLY, INSTALLATION AND COMMISSIONING OF INVERTED FLUORESCENCE MICROSCOPE.

Pre-Bid meeting was held on April 16th, 2021 at 16:30 Via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

21-04-2021

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF INVERTED FLUORESCENCE MICROSCOPE

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0308-20

DATE: 16/04/2021

	Query on existing specs	Explanation/modification
1	More details and type of Plan Achromat 4x (WD 10) Plan Achromat 20XC (WC 2)	Plan Fluorite objective 4x/5x with NA 0.13 or more, W.D 15 mm for Bright Field / Phase Contrast Purposes. Plan Fluorite objective 20 XC with NA 0.45 or more, W.D 8 to 6 mm or better
2	Plan Apo 60 oil /1.4 NA, W.D 0.13mm or better and DIC	Magnification 63 X or close to that is also is also possible
3	More details on Camera	10 megapixel or more Quantum efficiency more than 75% Sensor size: diagonal 17 mm or more. Pixel size 3.4 um bigger
4	Illumination	LED source only



**IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF INVERTED FLUORESCENCE MICROSCOPE**

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER IISERT-PUR-0308-20

DATE: 16/04/2021

Sr. No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL