



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH
TIRUPATI

CLARIFICATION ON TENDER NUMBER – Tender No: : IISERT/PUR/0324/19

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING FOUR PORT STANDALONE GLOVE BOX

Refer our Press Tender No .IISERT/PUR/0324/19 dated 29/11/2019 for procurement of Supply, Installation and Commissioning of Four Port Standalone Glove Box

Pre-Bid meeting was held on December 06th, 2019 at 14:30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

06.12.2019

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF FOUR PORT STANDALONE GLOVE BOX
TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0324/19

DATE: 06/12/2019

S.No	Query/Clarification Sought	Clarification / Amendment
1.	<u>Main Chamber:</u> -Service port feed through: 3 X DN40 ports + one USB feed through - Provision to upgrade to sliding shelves at any time, requested to remove due to our customization. - Free space on bottom side of the glove box	<u>Main Chamber:</u> -Service port feed through: 4 X DN40 ports (2 CF type, 2 thread type) + 2 X DN60 ports (1 CF type, 1 thread type) + Two USB feed through (in each two or three ports should be) - Sliding shelves option removed. - Bottom side of the glove box we need maximum free space to keep our thermal evaporator and vacuum pumps.
2.	<u>PLC Controller:</u> - Automatic Regeneration (With autorestart facility)	We have revised the specification based on their request We agreed to remove this specification.
3.	Two direct drive rotary vane vacuum pump to be decrease to one	One pump is enough - Since the second pump is an optional and single pump is capable of performing the required action based on their request we removed this one.
4.	Service experience of minimum 10 systems in research oriented institutes in India - request to decrease the number	Based on their request this has been changed as "At least 5 installations in research oriented institutes like IITs, IISERs, IISc, CSIR labs, NITs, Central universities etc."



ANNEXURE -III

IISER TIRUPATI

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF FOUR PORT STANDALONE GLOVE BOX
COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0324/19

DATE: 06/12/2019

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----