



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI
(An Autonomous Institute of National Importance under Ministry of Education, Govt. of India)
C/o. Sree Rama Engineering College, Rami Reddy Nagar, Karakambadi Road,
Mangalam P.O., Tirupati – 517 507, Andhra Pradesh.
Website: <http://www.iisertirupati.ac.in/>

CLARIFICATION ON TENDER NUMBER: IISERT-PUR-0465-21

ITEM DESCRIPTION- SUPPLY, INSTALLATION, AND COMMISSIONING OF TABLE TOP HRMS SYSTEM.

REFER OUR TENDER NO: IISERT-PUR-0465-21, DATED:23/03/2022 FOR SUPPLY, INSTALLATION, AND COMMISSIONING OF TABLE TOP HRMS SYSTEM.

The pre-Bid meeting was held on March 30th, 2022 at 15:30 Via Google Meet, and the minutes of the meeting are as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders briefed in general on the scope of the Project and thereafter requested Deputy Registrar (A&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure-II**

The representatives present were satisfied with the replies given and it was informed that the corrections/additions/clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with a vote of thanks to the Chair

04-04-2022

Sd/-
Deputy Registrar (A&P)



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

(An Autonomous Institute of National Importance under Ministry of Education, Govt. of India)

C/o. Sree Rama Engineering College, Rami Reddy Nagar, Karakambadi Road,

Mangalam P.O., Tirupati – 517 507, Andhra Pradesh.

Website: <http://www.iisertirupati.ac.in/>

Annexure: II

PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION, AND COMMISSIONING OF TABLETOP HRMS SYSTEM

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0465-21

DATE: 30/03/2022

Sr. No	Query/Clarification Sought	Clarification / Amendment
1	Chapter 4, Page no-34, point number 5 A vendor requested to specify the calibration type (internal or external) for achieving the specified ppm mass accuracy.	Chapter 4, Page no-34, point number 5 The specified mass accuracy is required against internal calibration. Mass accuracy for the external calibration can be tolerated up to 3 ppm in both MS and MS/MS modes.
2	Chapter 4, Page no-35, point number 10 One vendor sought clarity on the acquisition speed.	Chapter 4, Page no-35, point number 10 Please read it as ≥ 10 Hz scan speed.



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

(An Autonomous Institute of National Importance under Ministry of Education, Govt. of India)

C/o. Sree Rama Engineering College, Rami Reddy Nagar, Karakambadi Road,

Mangalam P.O., Tirupati – 517 507, Andhra Pradesh.

Website: <http://www.iisertirupati.ac.in/>

PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION, AND COMMISSIONING OF TABLETOP HRMS SYSTEM COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0465-21

DATE: 30/03/2022

Sr. No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL