



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0072/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF TABLE TOP NMR SPECTROMETER

REFER OUR TENDER NO: IISERT/PUR/0072/20 DATED 12-MAY-2020 FOR SUPPLY, INSTALLATION AND COMMISSIONING OF TABLE TOP NMR SPECTROMETER

Pre-Bid meeting was held on May 18th, 2020 at 14:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

18.05.2020

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI

PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF TABLE TOP NMR SPECTROMETER

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0072/20

DATE: 18/05/2020

Sr.No	Query/Clarification Sought	Clarification / Amendment
1.	The Technical specifications are modified according to following queries: Solvent suppression facility: The intense resonances of the solvents used need to be suppressed to maximize the sensitivity and spectra quality.	The following specifications are added additionally with the existing specifications as point no. 22 & 23. 22. Solvent suppression facility should be mandatorily available and is a part of the main technical specification which shall be counted for technical evaluation of the bid.
2.	Proof of supply and installation of the product quoted	23. The vendor should provide details of installation of the product quoted or purchase order copies of the same.



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF TABLE TOP NMR SPECTROMETER

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0072/20

DATE: 18/05/2020

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----