



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

Transit campus: C/o Sree Rama Engineering College Campus, Rami Reddy Nagar,
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: www.iisertirupati.ac.in

CORRIGENDUM: TECHNICAL BID

Tender Id	2022_ISRTP_712949_1
Tender Reference Number	IISERT/PUR/0022/22
Tender Title	Supply, Installation, and Commissioning of Fluorescence Spectrometer

With reference to the above tender, the Corrigendum has been made due to a change in the technical specification to provide opportunity for different vendors to participate.

Sr. No	Existing Specification	Amendment Specification
01	Time-resolved fluorescence	Instrument with or without time-resolved fluorescence is acceptable
02	Life time measurement	Instrument with or without life time measurement is acceptable
03	Wavelength Repeatability: 0.2 nm	Wavelength Repeatability: 0.3 nm or better
04	Light source: Pulsed Xenon Lamp	Pulsed Xenon Lamp or Xenon Arc Lamp
05	Five years replacement warranty on Lamp	An additional/spare lamp should be provide for durability instead of 5 year warranty
06	PMT detectors for both Excitation and emission detector	Excitation detector (PMT/Silicon photodiode) and emission detector (PMT/Silicon photodiode)
07	Sensitivity (RMS): 750:1 or better	Sensitivity (RMS): 2000:1 or better

All the Prospective Bidders are required to take cognizance of the proceedings of the document before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER Tirupati website <http://www.iisertirupati.ac.in> will remain unchanged.

**Sd/-
Deputy Registrar (Admin & Purchase)**