



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT/PUR/0698/18

ITEM DESCRIPTION- Supply, Installation and Commissioning of Dynamic Light Scattering with Zeta Potential

Refer our Press Tender Notice No.IISERT/S&P/07/19 dated 29.4.2019 for procurement of Dynamic Light Scattering with Zeta Potential.

Tender Reference Number – IISERT/PUR/0698/18

Pre-Bid meeting was held on May 06<sup>th</sup>, 2019 at 13.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

06.05.2019

Sd/-  
Assistant Registrar (S&P)

**IISER TIRUPATI**

**PRE-BID CONFERENCE FOR PROCUREMENT OF Dynamic Light Scattering with Zeta Potential**

**TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0698/18

DATE : 06.05.19

<b>S.No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
1	Disposable folded capillary cells Pack of 20	This has been changed to "Disposable capillary cells Pack of 20"
2	Standards - 1 No Zeta potential reference material & 1 No 220 nm	This has been changed to "Standards - 1 No Zeta potential reference material & 1 No $\leq$ 220 nm"
3	Multi-Angle Dynamic Light Scattering (with minimum 3 measurement angles, Goniometer type not acceptable).	No change in the minimum number of angles



ANNEXURE -III

**IISER TIRUPATI**

**PRE-BID CONFERENCE FOR PROCUREMENT OF Dynamic Light Scattering with Zeta Potential**

**COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0698/18

DATE : 06.05.19

<b>S.No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
	-----NIL-----	-----NIL-----