



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

### TIRUPATI

#### CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0062-18

#### ITEM DESCRIPTION- PROCUREMENT OF UPRIGHT RESEARCH GRADE POLARIZING MICROSCOPE

Refer our Press Tender Notice No. IISERT/S&P/03/18 dated 21.06.2018 for procurement of UPRIGHT RESEARCH GRADE POLARIZING MICROSCOPE. Tender Reference Number - IISERT-PUR-0062-18.

Pre-Bid meeting was held on June 28<sup>th</sup>, 2018 at 14.30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iisertirupati.ac.in](http://www.iisertirupati.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

12.07.2018

Sd/-  
Assistant Registrar (S&P)



ANNEXURE -II

**IISER TIRUPATI**  
**PRE-BID CONFERENCE FOR PROCUREMENT OF UPRIGHT RESEARCH GRADE POLARIZING MICROSCOPE**  
**TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - **IISERT-PUR-0062-18**

DATE: 28.06.18

S.No	Query/Clarification Sought	Clarification/Amendment
01.	<p><b>Page No. 22, Point. 12</b></p> <p>A vendor suggested that the motorized condenser with motorized top lens, automatic Koehler illumination should be removed as we are asking for manual microscope not motorized one.</p>	<p><b>Page No. 22, Point. 12</b></p> <p>Agreed and motorized condenser with motorized top lens, automatic Koehler illumination is removed from the specification.</p>
02.	<p><b>Page No. 23, Point 16c</b></p> <p>A vendors requested to change the suggested to reduce the minimum working distance for 10x objective should be kept at 17.5 mm or better.</p> <p>Rest three vendors agreed with the specification.</p>	<p><b>Page No. 23, Point 16c</b></p> <p>Agreed and the minimum working distance now modified as 17.5 mm or better for 10x objective.</p>
03.	<p><b>Page No. 23, Point 17g</b></p> <p>A vendor requested to change the minimum pixel size <math>3.2 \times 3.2 \mu\text{m}</math> <math>2.35 - 7 \mu\text{m}</math> or superior considering the maximum image size quoted against point 17d.</p>	<p><b>Page No. 23, Point 17g</b></p> <p>Agreed and has been changed to minimum pixel size <math>2.35 - 7 \mu\text{m}</math> or superior.</p>



ANNEXURE -III

IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF UPRIGHT RESEARCH GRADE POLARIZING MICROSCOPE**

**COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0062-18

DATE: 28.06.18

S.No	Query/Clarification Sought	Clarification / Amendment
01	Nil	Nil