



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0425-17

ITEM DESCRIPTION- Glove Box Workstation

Refer our Press Tender Notice No.IISERT/S&P/09/17 dated 12.01.2018 for procurement of Glove Box Workstation. Tender Reference Number - IISERT-PUR-0425-17.

Pre-Bid meeting was held on January 18th, 2018 at 15.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iisertirupati.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

29.1.2018

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF GLOVE BOX WORKSTATION
 TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0425-17

DATE: 29.01.18

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page No. 22 Inner Box Dimensions 900-950 mm x 1250-1300 mm x 775-800 mm (height x length x diameter)	Modified to, Inner Box Dimensions 900-950 mm x 1200-1300 mm x 750-800 mm (height x length x diameter)
2	Page No. 22 2 Polymer glove ports, 210-240 mm diameter, should be O ring sealed	Modified to, 2 Polymer glove ports, 200-240 mm diameter, should be O ring sealed
3	Page No. 22 Stainless steel Cylindrical antechamber (380 to 400 mm diameter, Length 590 to 610 mm) with SS sliding tray (at the right hand side of the box)	Modified to, Stainless steel Cylindrical antechamber (370 to 400 mm diameter, Length 590 to 610 mm) with SS sliding tray (at the right hand side of the box)
4	Page No. 23 Is the Glove box workstation is modular?	Clarification: Yes. The glovebox should be modular and expandable in future upgradation



IISER TIRUPATI

**PRE-BID CONFERENCE FOR PROCUREMENT OF GLOVE BOX WORKSTATION
COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT-PUR-0425-17

DATE: 29.01.18

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----