

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0384-17

ITEM DESCRIPTION – **UV-Vis-NIR Spectrophotometer**Refer our OPEN TENDER NOTICE NO: 08/2017 dated 15.12.2017 for procurement of **UV-Vis-NIR Spectrophotometer**.

Tender Reference Number – IISERT-PUR-0384-17.

Pre-Bid meeting was held on December 22nd, 2017 from 14:30-16:30 at IISER Tirupati and the Minutes of Meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure –II.

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iisertirupati.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

Sd/-Assistant Registrar (S&P)

05.01.2018



IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF UV-Vis NIR Spectrophotometer TECHNICAL QUERIES AND CLARIFICATION

DATE: 05.01.18

TENDER NUMBER - IISERT-PUR-0384-17

S.No	Query/Clarification Sought	Clarification / Amendment
1	Page no. 22, Chapter 4, Maximum scan rate We request you to kindly amend the maximum scan rate for NIR as 5000 nm/min	Specifications for maximum scan rates are amended as: Maximum scan rate (nm/min):2000 (UV-vis.), 5000 (NIR)
2	Page no. 23, Chapter 4, Cuvettes Specification	Specifications for maximum scan rates are amended as:
	Our system comes with Cuvettes capable of measuring absorption in the range 200-3300 nm; 10 mm, 3.5 mL volume; with 10mm path length, volume-1.4 ml \pm 0.1 mL; capable of measuring absorption in the range 200-3300 nm; 10 mm, 400 μL volume and capable of measuring absorption in the range 200-3300 nm; 10 mm, 50 μL volume. So, we request you to kindly amend the specification to 50 μL volume and 400 μL volume for the cuvettes.	One pair, 10mm path length, volume-1.4 ml \pm 0.1 mL. One pair of cuvettes capable of measuring absorption in the range 200-3300 nm; 10 mm, 3.5 mL volume. One pair of cuvettes capable of measuring absorption in the range 200-3300 nm; 10 mm, 50 μ L volume. One pair of cuvettes capable of measuring absorption in the range 200-3300 nm; 10 mm, 400 μ L volume
3	Page no. 23, Chapter 4, Software	Vendors are required to supply:
	Our system does not come with the computer and printer. So, we request you to kindly amend the specification with additional requirements of one Printer with one Branded Desktop Computer /	One Printer with one Branded Desktop Computer / Laptop with Core i5 Processor, 4GB RAM, 1TB HDD along with the



	Laptop with Core i5 Processor, 4GB RAM, 1TB HDD should be provided with the system	system
4	Page no. 23, Chapter 4, Point24.	Vendors are required to supply:
	Our system does not come with the water circulation system with chiller.	Complete water circulation set up with chiller along with the system.
	So, we request you to kindly amend the specification with additional requirements of Complete water circulation set up with chiller should be provided with the system.	

DATE: 05.01.18



IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF UV-Vis NIR Spectrophotometer COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0384-17

S.No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL