



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0345-17

ITEM DESCRIPTION- PROCUREMENT OF HIGH-RESOLUTION TISSUE SCANNER

Refer our Press Tender Notice No.IISERT/S&P/08/17 dated 14.12.2017 for procurement of HIGH-RESOLUTION TISSUE SCANNER. Tender Reference Number - IISERT-PUR-0345-17.

Pre-Bid meeting was held on December 22nd , 2017 at 16.00 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website [www.iisertirupati.ac .in](http://www.iisertirupati.ac.in) will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

04.01.2018

Sd/-
Assistant Registrar (S&P)

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH-RESOLUTION TISSUE SCANNER

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT-PUR-0345-17

DATE: 04.01.18

S.No	Query/Clarification Sought	Clarification/Amendment
01.	<p>Page No. 22, Chapter 4</p> <p><i>Specification number 12.</i> A vendor mentioned that each image data size is likely to be high (around a GB) and therefore enough space and backup is required in the computer.</p>	<p>Page No. 22, Chapter 4</p> <p>The tender specification is modified as below: Please read ≥ 4 TB SSD is required.</p>
02.	<p>Page No. 22, Chapter 4</p> <p><i>Specification number 11.</i> A vendor pointed out that for the safe performance of the instrument an UPS is required for avoiding the disturbance caused by power fluctuation, if happens.</p>	<p>Page No. 22, Chapter 4</p> <p>An appropriate UPS for running the instrument is accepted.</p>

**IISER TIRUPATI****PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH-RESOLUTION TISSUE SCANNER****COMMERCIAL QUERIES AND CLARIFICATION**TENDER NUMBER - **IISERT-PUR-0345-17**

DATE: 04.01.18

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----