



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER - IISERT-PUR-0224-16

ITEM DESCRIPTION- HIGH SPEED CENTRIFUGE

Refer our Press Tender Notice No.IISERT/S&P/03/17-18 dated 12.08.2017 for procurement of High Speed centrifuge. Tender Reference Number - IISERT-PUR-0224-16.

Pre-Bid meeting was held on August 23rd, 2017 at 14.30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iisertirupati.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

30.8.2017

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH SPEED CENTRIFUGE

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0224-16

DATE : 23.08.17

S.No	Query/Clarification Sought	Clarification / Amendment
01	Requested to remove the word aluminum in all rotor material description for all rotors. OR Rotor Material description in Each Rotor Asked to Aluminium (or) Carbon Fiber (or) Titanium.	No change in specifications
02	Request to change the No.of Programs to 100.	Amended as requested
03	Query if the rotor with 1000ml capacity can be removed	No change in specifications



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR PROCUREMENT OF HIGH SPEED CENTRIFUGE

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISER-PUR-0224-16

DATE : 23.08.17

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----