



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

ransit campus: C/o Sree Rama Engineering College Campus, Rami Reddy Nagar,
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: www.iisertirupati.ac.in

Corrigendum: Technical Bid

<u>Tender Id</u>	2019_ISRTP_506054_1
<u>Tender Reference Number</u>	IISERT/PUR/0372/19
<u>Tender Title</u>	Supply Installation and Commissioning of Mass spectrometer

With reference to the above tender the Corrigendum has been made due to change in technical specification.

Sr.No	Point No	Existing Specification	Amendment Specification
1	4	Added in Section 1.1. as point no 10: "Sensitivity: For 10pg or less reserpine, S/N ratio should be better than 100:1."	Added in Section 1.1. as point no 10: "Sensitivity: For 10pg or less reserpine, S/N ratio should be 100:1 or better."

All the Prospective Bidders are required to take cognizance of the addendum along with proceedings of the Corrigendum before submitting their bids as stipulated in the Bidding document

Assistant Registrar