

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0060/19

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF STOPPED FLOW MODULE FOR THE EXISTING JASCO CD SPECTROMETER

Tender Reference Number -IISERT/PUR/0060/19.dated 10/06/2019 for Supply, Installation and Commissioning of Stopped Flow Module for the Existing JASCO CD Spectrometer.

Pre-Bid meeting was held on June 14th, 2019 at 15.30 and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additons / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iisertirupati.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

DATE: 14.06.19

Sd/-
Assistant Registrar (S&P)



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

ANNEXURE -I

PRE-BID CONFERENCE FOR PROCUREMENT OF SUPPLY, INSTALLATION AND COMMISSIONING OF STOPPED FLOW MODULE FOR THE EXISTING JASCO CD SPECTROMETER

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER -: IISERT/PUR/0060/19

DATE: 14.06.19

S. No	Query/Clarification Sought	Clarification / Amendment
1.	The vendor sought a clarification on the priming volume. The existing specification is: 2(b)priming volume: 88 µl/syringe or better	Existing Tender specifications prevail. The following point has been added: Priming/Dead volume: 88 µl/syringe or better
2.	The vendor sought a clarification on the path length. The existing specification is: 3 (a) light path: 1.5mm and 0.75 mm	The specifications has been modified to: 3 (a) light path range : 0.5mm - 2 mm



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

ANNEXURE -II

IISER TIRUPATI

PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION AND COMMISSIONING OF STOPPED FLOW MODULE FOR THE EXISTING JASCO CD SPECTROMETER

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER -: IISERT/PUR/0060/19

DATE : 14.06.19

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----