



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI**

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar,  
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: <http://www.iisertirupati.ac.in/>

**CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0465/21**

**ITEM DESCRIPTION-** SUPPLY, INSTALLATION, AND COMMISSIONING OF TABLE TOP HRMS SYSTEM.

REFER OUR TENDER NO: IISERT/PUR/0465/21, DATED:07/12/2021 FOR SUPPLY, INSTALLATION AND COMMISSIONING OF TABLE TOP HRMS SYSTEM.

Pre-Bid meeting was held on December 13<sup>th</sup>, 2021 at 15:30 Via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders briefed in general the scope of the Project and thereafter requested Deputy Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

16.12.2021

Sd/-  
**Deputy Registrar (A&P)**



## भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI**

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar,  
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: <http://www.iisertirupati.ac.in/>

**Annexure II**

### **PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION AND COMMISSIONING OF TABLE TOP HRMS SYSTEM**

#### **TECHNICAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0465/21**

**DATE: 14/12/2021**

<b>Sr. No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
1	<b>Chapter 4, Page no-24, point number 4</b> A vendor asked for the information on the desired m/z range for the specified resolution.	<b>Chapter 4, Page no-24, point number 4</b> Please read the m/z range 50-1000 for $\geq 60,000$ resolution
2	<b>Chapter 4, Page no-24, point number 5</b> A vendor requested to specify the calibration type (internal or external) for achieving the specified ppm mass accuracy.	<b>Chapter 4, Page no-24, point number 5</b> The specified mass accuracy is required against both internal and external calibration.
3	<b>Chapter 4, Page no-24, point number 10</b> One vendor sought clarity on the acquisition speed.	<b>Chapter 4, Page no-24, point number 10</b> Please read it as $\geq 10$ Hz scan speed.
4	<b>Chapter 4, Page no-26, point number 18f</b>	<b>Chapter 4, Page no-26, point number 18f</b> Please read the backup for $\geq 2$ h.



## भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI**

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar,  
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: <http://www.iisertirupati.ac.in/>

	One vendor requested to reduced the battery backup to 1h – as the total space required for this bakup would be too bulky otherwise.	
5	<b>Chapter 4, Page no-26, point number j)</b> A vendor requested to change the PDA detector range up to 800 nm	<b>Chapter 4, Page no-26, point number j)</b> This can be considered also.
6	<b>Chapter 4, Page no-26, point number m)</b> A vendor requested to change the column heater temperature range from 4-80 °C.	<b>Chapter 4, Page no-26, point number m)</b> This can be considered also.



भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI**

Transit campus: Sree Rama Engineering College Campus, Rami Reddy Nagar,

Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: <http://www.iisertirupati.ac.in/>

**Annexure III**

**PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION AND COMMISSIONING OF TABLE TOP HRMS SYSTEM**

**COMMERCIAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0465/21**

**DATE: 14/12/2021**

<b>Sr. No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
01	Price Schedule Form As per form it appears that we have to submit our offer in Indian Currency. Please clarify whether to offer in Indian currency or US Dollars	The tender is released for Domestic Manufacturers only, hence the offer should be in Indian Currency only.