

#### INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

#### CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0014/21

#### ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF MALDI TOF/TOF SYSTEM.

REFER OUR TENDER NO: IISERT/PUR/0014/21, DATED:21/05/2021 FOR SUPPLY, INSTALLATION AND COMMISSIONING OF MALDI TOF/TOF SYSTEM.

Pre-Bid meeting was held on May 31<sup>st</sup>, 2021 at 15:30 Via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II** 

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website http://www.iisertirupati.ac.in/ will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

01.06.2021

Sd/-Assistant Registrar (S&P)



**Annexure II** 

#### IISER TIRUPATI PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION AND COMMISSIONING OF MALDI TOF/TOF SYSTEM

## **TECHNICAL QUERIES AND CLARIFICATION**

# TENDER NUMBER - IISERT/PUR/0014/21

DATE: 31/05/2021

Sr. No	Query/Clarification Sought	Clarification / Amendment
	Chapter 4, Page no-25, point number 4	Chapter 4, Page no-25, point number 4
1	A vendor asked for the information on external calibration of the instrument.	We need $\leq$ 10 ppm accuracy with external calibration.



## IISER TIRUPATI PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF MALDI TOF/TOF SYSTEM

## **COMMERCIAL QUERIES AND CLARIFICATION**

## **TENDER NUMBER - IISERT/PUR/0014/21**

DATE: 31/05/2021

Sr. No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL