



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

### CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0048/21

#### ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY.

REFER OUR TENDER NO: IISERT/PUR/0048/21 DATED:10/05/2021 FOR SUPPLY, INSTALLATION AND COMMISSIONING OF HPLC.

Pre-Bid meeting was held on May 15<sup>th</sup>, 2021 at 11:00 Hrs. Via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

15-05-2021

Sd/-  
Assistant Registrar (S&P)



**Annexure II**

**IISER TIRUPATI  
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY**

**TECHNICAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0048/21**

**DATE: 15/05/2021**

<b>Sr. No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
1	<b>Eligibility Criteria</b> Technical specification: Autosampler capacity:	Minimum of 70 vials (1.5 mL) or above.  Deleted 36 vials of 4 ML.
2	Column Oven: Temperature precision: +/- 0.05 ° C	Change to +/- 0.1 ° C
3	Temperature range: Previous - from 10°C below ambient to 85°C	Change to from 10°C below ambient to 65°C or above.

4.	<p>Detector Type:</p> <p>Previous: 1024-element diode array</p>	<p>Change to: 512 / 1024-element diode array</p>
5.	<p>Detector wavelength range: previous - Wavelength range must be from 190-900nm or better</p>	<p>Change to Wavelength range must be from 190-800nm or better</p>
6	<p>Detector sampling rate: Previous – 80 Hz or above.</p>	<p>Change to 65 Hz or above</p>
7	<p>Fraction collector: system flow rate – previous: 10 ml/min</p>	<p>Change to 5 mL / min.</p>
8	<p>Fraction collector: Temperature controller for fraction collector</p>	<p>Change to Optional temperature controller for fraction collector.</p>



**Annexure III**

**IISER TIRUPATI**

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY  
COMMERCIAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0048/21**

**DATE: 15/05/2021**

<b>Sr. No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
	NIL	NIL