



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

### CLARIFICATION ON TENDER NUMBER: **IISERT/PUR/0399/20**

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY.

REFER OUR TENDER NO: **IISERT/PUR/0399/20** DATED 19-MARCH-2021 FOR PROCUREMENT OF SUPPLY, INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY.

Pre-Bid meeting was held on March 25<sup>th</sup> 2021 at 14:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

05.04.2021

Sd/-  
**Assistant Registrar (S&P)**

IISER TIRUPATI  
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY

**TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0399/20

DATE: 05/04/2021

S.No	Query/Clarification Sought	Clarification / Amendment
1	Active Inlet Valve option	Active inlet valve option or its equivalent.
2	Hydraulic system	Dual pistons in series / parallel for higher robustness.
3	Capacity	minimum of 70 vials (1.5 mL) or 36 vials (4 mL) or equivalent or above.
4	Independent temperature zones	At least one or more independent temperature zones
5	Sampling rate	Minimum of 100 Hz or above
6	Trays and Tubes	Must be suitable with vials, test tubes etc.
7	Suitable Accessories Required	<ol style="list-style-type: none"><li>1. Solvent Filtration Unit</li><li>2. Sonicator unit</li><li>3. PC with monitor of appropriate specification for robust data analyses</li><li>4. UPS with a 3 KVa power back up up to 30 minutes of power supply.</li></ol>

**IISER TIRUPATI**

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF HIGH-PERFORMANCE LIQUID CHROMATOGRAPHY**

**COMMERCIAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0399/20**

**DATE: 05/04/2021**

<b>S.No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
	-----NIL-----	-----NIL-----