



CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0182/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF FLUORIMETER.

Refer our Tender No: IISERT/PUR/0182/20 dated 29th-Dec-2020 Supply, Installation and Commissioning of Fluorimeter.

Pre-Bid meeting was held on Jan 06th, 2021 at 15:30 Via Google Meets and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

06/01/2021

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF FLUORIMETER
TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0182/20

DATE: 06/01/2021

Sr.No	Query/Clarification Sought	Clarification / Amendment
1.	<u>Wavelength accuracy: ± 0.1 nm or better.</u> Can you change the specification to ± 0.5 nm or better?	-We have agreed to change the wavelength accuracy to ± 0.5 nm or better as per the request
2	<u>Water Raman (S/N): Sensitivity specification: $>20,000:1$.</u> Can you change the specification to $>15,000:1$ or more	We have agreed to change the Water Raman (S/N): Sensitivity specification to $>15,000:1$ (R.M.S method) or more as per the request
3	<u>Peltier specification: -25 to 105°C.</u> Can you change the temperature range?	We have agreed to change the temperature range to -15 to 105°C .
4	<u>Table to mount the spectrometer and a cupboard to store the accessories.</u> Please let us know the size of the table.	The vendors for the spectrometer were asked to provide their instrument dimensions to their local suppliers for fetching the configuration of the table



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF FLUORIMETER
COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0182/20

DATE: 06/01/2021

Sr.No	Query/Clarification Sought	Clarification / Amendment
01	<p>Page No:20 Point No 16.3 IISER, TIRUPATI reserves the right to cancel the order in case the delay is more than 0 weeks. Penalties, if any, will be deducted from the Security Deposit</p>	<p>Page No:20 Point No 16.3 IISER, TIRUPATI reserves the right to cancel the order in case the delay is more than 10 weeks. Penalties, if any, will be deducted from the Security Deposit</p>