



# भारतीय विज्ञान शिक्षा एवं अनुसंधान संस्थान तिरुपति

**INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI**

Transit campus: C/o Sree Rama Engineering College Campus, Rami Reddy Nagar,  
Karakambadi Road, Mangalam (B.O), Tirupati - 517 507

Website: [www.iisertirupati.ac.in](http://www.iisertirupati.ac.in)

## **CORRIGENDUM: TECHNICAL BID**

<b>Tender Id</b>	2020_ISRTP_561340_1
<b>Tender Reference Number</b>	IISERT/PUR/0034/20
<b>Tender Title</b>	Supply, Installation and Commissioning of Microvolume UV-Vis Spectrophotometer at IISER Tirupati Main Campus, Yerpedu

With reference to the above tender the Corrigendum has been made due to change in technical specification.

Sr.No	Point No	Existing Specification	Amendment Specification
1	#1	Wavelength Range: 198 to 1000nm	Wavelength Range: 200 to 650nm or more
2	#10	dsDNA Maximum Concentration: (0.2mm): 6000 ng/ $\mu$ l	dsDNA Maximum Concentration: (0.2mm): 7500 ng/ $\mu$ l or more

All the Prospective Bidders are required to take cognizance of the proceedings of the document before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER Tirupati website <http://www.iisertirupati.ac.in> will remain unchanged.

**Assistant Registrar (Admin & Purchase)**