

INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0049/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF FLASH CHROMATOGRAPHY

REFER OUR TENDER NO: IISERT/PUR/0049/20 DATED 1-MAY-2020 FOR SUPPLY, INSTALLATION AND COMMISSIONING OF FLASH CHROMATOGRAPHY

Pre-Bid meeting was held on May 11th, 2020 at 14:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website http://www.iisertirupati.ac.in/ will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

11.05.2020

Sd/-Assistant Registrar (S&P)



IISER TIRUPATI PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF FLASH CHROMATOGRAPHY <u>TECHNICAL QUERIES AND CLARIFICATION</u>

TENDER NUMBER - IISERT/PUR/0049/20

DATE: 11/05/2020

S.No	Query/Clarification Sought	Clarification / Amendment
	The Technical specifications are modified according to following:	New Corrected Text:
1.	• Pumping system with maximum pressure capacity of 200 psi or more with flow Rate of 1 to100ml/min or more	• Pumping system with <u>minimum</u> pressure capacity of 200 psi or more with flow Rate of 1 to100ml/min or more.
2.	• 30 numbers of 4g, 50 numbers of 12g, 20 numbers of 24g, 20 numbers of 40g and 20 numbers of 100 g and 10 numbers of 350 g prepacked refillable columns and 5 numbers of C18 columns of different sample loading capacity to be supplied along with the system.	• 30 numbers of columns comprise of <u>minimum of</u> 4g, 50 numbers with a <u>minimum of</u> 12g, 20 numbers with a <u>minimum of</u> 24g, 20 numbers with a <u>minimum of</u> 40g and 20 numbers with a <u>minimum of</u> 100 g and 10 numbers with a <u>minimum of</u> 350 g prepacked refillable columns and 5 numbers of C18 columns of different sample loading capacity to be supplied along with the system. <u>Equivalent specs to these quantities are acceptable.</u>
3.	• Safety features like automatic stop of flow if collection tray is full or unavailable, leak detection, solvent level detection, pressure moderating system.	• Criteria of leak detection feature may be considered <u>OPTIONAL</u> .



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF FLASH CHROMATOGRAPHY COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0049/20

DATE: 11/05/2020

Sr.No	Query/Clarification Sought	Clarification / Amendment
	NIL	NIL