



## INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0060/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF UV-VIS-WITH-SPECTRO-ELECTROCHEMICAL ATTACHMENT

Refer our Tender No: IISERT/PUR/0060/20 dated 29-April-2020 for procurement of Supply, Installation and Commissioning of UV-Vis-with-Spectro-electrochemical attachment

Pre-Bid meeting was held on May 06<sup>th</sup>, 2020 at 15:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

06.05.2020

Sd/-  
Assistant Registrar (S&P)



IISER TIRUPATI

PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF UV-VIS-WITH-SPECTRO-ELECTROCHEMICAL ATTACHMENT

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0060/20

DATE: 06/05/2020

S. No	Query/Clarification Sought	Clarification / Amendment
1.	Clarification required: Should we consider output voltage or Applied Potential. If we need to consider Applied potential, we believe, it should be +/- 20V.	- As far as electrochemistry design is concerned, compliance is control voltage. So Instead of Maximum applied potential: 20V or better <b>We will modify the applied voltage to +/-10V</b>
2.	<b><u>About Current range and resolution</u></b> Requested to change the current resolution	It should be in pA on the lowest current range or better
3.	<b><u>Numbers of Uv-Vis-With-Spectro-Electrochemical and computers</u></b>	Two numbers of UV-Vis-With-Spectro-Electrochemical instruments with two sets of computers



**IISER TIRUPATI**

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF UV-VIS-WITH-SPECTRO-ELECTROCHEMICAL ATTACHMENT**

**COMMERCIAL QUERIES AND CLARIFICATION**

**TENDER NUMBER - IISERT/PUR/0060/20**

**DATE: 06/05/2020**

<b>S.No</b>	<b>Query/Clarification Sought</b>	<b>Clarification / Amendment</b>
	-----NIL-----	-----NIL-----