



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0989/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF STAINLESS STEEL CHAMBER FOR THERMAL EVAPORATION WITH SUPPORT STAND

Refer our Tender No: IISERT/PUR/0989/20 dated 05th April 2021 for SUPPLY, INSTALLATION AND COMMISSIONING OF STAINLESS STEEL CHAMBER FOR THERMAL EVAPORATION WITH SUPPORT STAND

Pre-Bid meeting was held on April 12th, 2021 at 16:30 Via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

20-04-2021

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION AND COMMISSIONING OF STAINLESS STEEL CHAMBER FOR THERMAL EVAPORATION WITH SUPPORT STAND

TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0989/20

DATE: 12/04/2021

S.No	Query/Clarification Sought	Clarification / Amendment
1.	Number of K-Cells Tender mentions a cluster flange having 5 Nos Knudson cells as mentioned under para 1 CHAMBER point 2. But later under para 4 only one Knudson Cell with dc power supply unit is described. It is not clear if (a) You want 5 numbers K-Cells, (b) You want provision to mount 5 K-Cells but only one should be supplied at the moment.	We require provision for 5 Nos of Knudson cells with proper sealing in the chamber. But currently we require only one K-cell for metal evaporation.
2.	Flange Size Information on the vacuum pumping system (supplied free issue) must be given as we must design and fabricate a process chamber compatible with the flange of vacuum pump.	Turbo molecular pump flange size is DN 160 CF-F and for rotary van pump the flange size should be DN 16 ISO-KF
3.	Vacuum Gauges Information on vacuum gauges (supplied free issue) must be given in order to make flanges on the process chamber compatible.	The details are as follows: (a) Active Pirani/Cold Cathode Transmitter DN 25 ISO-KF
4.	Important Note: Suppliers are requested to provide product number, make (for both local/imported) and clear line drawings and brochure of all the major items in the tender document.	



ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR SUPPLY, INSTALLATION AND COMMISSIONING OF STAINLESS STEEL CHAMBER FOR THERMAL EVAPORATION WITH SUPPORT STAND

COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0989/20

DATE: 12/04/2021

Sr. No	Query/Clarification Sought	Clarification / Amendment
1	NIL	NIL