



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH

TIRUPATI

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0173/20

ITEM DESCRIPTION- SUPPLY, INSTALLATION AND COMMISSIONING OF CYCLIC VOLTAMMETRY

Refer our Tender No: IISERT/PUR/0173/20 dated 29-Sept-2020 for procurement of Supply, Installation and Commissioning of Cyclic Voltammetry

Pre-Bid meeting was held on Oct 05th, 2020 at 15:30 via Google Meet and minutes of meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Assistant Registrar (S&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure -II**

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

05.10.2020

Sd/-
Assistant Registrar (S&P)



IISER TIRUPATI
PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF CYCLIC VOLTAMMETRY
TECHNICAL QUERIES AND CLARIFICATION

TENDER NUMBER - IISERT/PUR/0173/20

DATE: 05/10/2020

Sr.No	Query/Clarification Sought	Clarification / Amendment
1.	Computers and Printer Printers	3 computers + 1 computer (central data keeping facility) = total 4 computers 4 printers
2.	Electrodes	Following Electrode should be added with instruments Rotating ring disc electrode (total 3 different setup for 3 instruments) Rotating ring electrode (total 3 different setup for 3 instruments)
3.	Impedance accuracy	No Change
4.	Booster	No Change



ANNEXURE -III

IISER TIRUPATI

**PRE-BID CONFERENCE FOR INSTALLATION AND COMMISSIONING OF CYCLIC VOLTAMMETRY
COMMERCIAL QUERIES AND CLARIFICATION**

TENDER NUMBER - IISERT/PUR/0173/20

DATE: 05/10/2020

S.No	Query/Clarification Sought	Clarification / Amendment
	-----NIL-----	-----NIL-----