

CLARIFICATION ON TENDER NUMBER: IISERT/PUR/0008/23

ITEM DESCRIPTION- UPGRADATION AND EXTENSION OF THE EXISTING INERT GAS WORKSTATION

Tender Reference Number –IISERT/PUR/0008/23 dated 15/04/2022 for Upgradation and Extension of the Existing Inert Gas Workstation.

Pre-Bid meeting was held on April 25th, 2023 at 15.30 Hrs. via Google Meet and the minutes of the meeting is as under.

At the outset, the Chairman welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter requested Deputy Registrar (A&P) to brief the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in Annexure -II

The representatives present were satisfied with the replies given and it was informed that the corrections / additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website www.iisertirupati.ac.in will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with vote of thanks to the Chair

DATE: 29.04.2023

Sd/-
Deputy Registrar (A&P)

**PRE-BID CONFERENCE FOR UPGRADATION AND EXTENSION OF THE EXISTING INERT GAS WORKSTATION
TECHNICAL QUERIES AND CLARIFICATION**

TENDER NUMBER -: IISERT/PUR/0008/23

DATE: 25.04.2023

| S. No | Query/Clarification Sought | Clarification / Amendment |
|-------|---|---|
| 1. | Equipped with vacuum hot-oven features operating between room temperature to 150 °C, PID controlled, uniform heating inside the antechamber, PLC controlled temperature and timer settings. (<i>Negotiable</i>) | This point has been removed |
| 2 | Refrigerator -40°C integrated with the system (<i>Negotiable based on the budget</i>) | Refrigerator -35 °C integrated with the system (<i>Negotiable based on the budget</i>) |



INDIAN INSTITUTE OF SCIENCE EDUCATION AND RESEARCH TIRUPATI

ANNEXURE -III

IISER TIRUPATI

PRE-BID CONFERENCE FOR UPGRADATION AND EXTENSION OF THE EXISTING INERT GAS WORKSTATION
COMMERCIAL QUERIES AND CLARIFICATION

TENDER NUMBER -: IISERT/PUR/0008/23

DATE: 25.04.2023

| S.No | Query/Clarification Sought | Clarification / Amendment |
|------|----------------------------|---------------------------|
| | | |
| | -----NIL----- | -----NIL----- |
| | | |
| | | |
| | | |