

ITEM DESCRIPTION-SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF BROADBAND SEISMOMETER.

Refer our Bid No: IISERT/PUR/0407/25 dated 13th January 2026 for Supply, Installation, Testing and Commissioning of Broadband Seismometer.

The Pre-Bid meeting was held on 16th January, 2026 at 11:30 Via Google Meet and the minutes of the meeting are as under.

At the outset, the Deputy Registrar (Academics and Purchase) welcomed all the Members and the representative of the Prospective Bidders and briefed in general the scope of the Project and thereafter briefed the vendors on the salient features of the commercial terms and the indenting Officer to read out the clarification sought by the Prospective Bidders and replied thereto as detailed in **Annexure-I**

The representatives present were satisfied with the replies given and it was informed that the corrections/additions / clarifications given, as discussed during the Pre-Bid Conference would be hosted on the website of IISER Tirupati and all the Prospective Bidders are required to take cognizance of the proceedings of the Pre-Bid Conference before submitting their bids as stipulated in the Bidding Documents.

The other terms & conditions of the notice issued on our IISER website <http://www.iisertirupati.ac.in/> and <https://eprocure.gov.in/eprocure/app> will remain unchanged. No more correspondence in this regard will be entertained

The meeting ended with a vote of thanks.

16/01/2026

Sd/-
Deputy Registrar (A&P)

Annexure - I

**PRE-BID CONFERENCE FOR - SUPPLY, INSTALLATION, TESTING AND
COMMISSIONING OF BROADBAND SEISMOMETER.**

TECHNICAL QUERIES AND CLARIFICATION

Tender No: IISERT/PUR/0407/25

Date: 13/01/2026

Sr. No	Query/Clarification Sought	Clarification / Amendment
1	One vendor requested for the relaxation of internal space (128 GB) of memory card.	The minimum supported capacity is 64 GB.

Apart from the one point mentioned above the remaining specifications are same as mentioned before.



PRE-BID CONFERENCE FOR - SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF BROADBAND SEISMOMETER.

COMMERCIAL QUERIES AND CLARIFICATION

Tender No: IISERT/PUR/0407/25

Date:13/01/2026

Sr. No	Query/Clarification Sought	Clarification / Amendment
1	The agency requested an increase in the delivery period for the item.	The Committee has considered the request and approved the amendment of the delivery period to 75 days in place of the earlier 60 days.
2	Tender Clause 16(16.2)- Penalty for delayed Services / LD: If the supplier fails to Supply, Install and Commission the system as per specifications mentioned in the order within the due date, the Supplier is liable to pay liquidated damages of 1% of order value per every week of delay subject to a maximum of 10% beyond the due date. Such money will be deducted from any amount due or which may become due to the supplier.	To be read as: If the supplier fails to Supply, Install and Commission the system as per specifications mentioned in the order within the due date, the Supplier is liable to pay liquidated damages of 0.5% of order value per every week of delay subject to a maximum of 10% beyond the due date. Such money will be deducted from any amount due or which may become due to the supplier.

